

**PARTISOL PARTICULATE MATTER (PM) SAMPLER
Calibration Data**

SITE NAME: KCBX ST-NE
AQS Site Code: N/A
DATE: 11-Mar-14
TIME: 14:00

CLIENT: Koch
SAMPLER ID: ST-NE-Metals
Model Number: 2025i
Serial Number: 20547

TEMPERATURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Sensor Temperature (T _s) (°C)	Calibration Sensor Temperature (T _c) (°C)	Temperature Difference ¹ (Sampler - Calibration) (°C)
	4.8	5.8	-1.0
	4.8	5.8	-1.0
	4.8	5.8	-1.0

1. Acceptance Criteria: ± 2.0 °C

PRESSURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Sensor Pressure (P _s) (mm Hg)	Calibration Sensor Pressure (P _c) (mm Hg)	Pressure Difference ² (Sampler - Calibration) (mm Hg)
	736.0	737.0	-1.0
	736.0	737.0	-1.0
	736.0	737.0	-1.0

2. Acceptance Criteria: ± 10.0 mm Hg

COMPARTMENT TEMPERATURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Compartment Temperature (T _s) (°C)	Calibration Sensor Temperature (T _c) (°C)	Temperature Difference ³ (Sampler - Calibration) (°C)
	6.7	5.9	0.8
	6.7	5.9	0.8
	6.7	5.9	0.8

3. Acceptance Criteria: ± 2.0 °C

LEAK CHECK, TIME, AND FLOWRATE Calibration Data				
External Leak Check ⁴ <input type="text" value="21.0"/> mmHg 4. Acceptance Criteria < 25 mm Hg		Clock/Timer Verification ⁵ : Partisol : <input type="text" value="14:13:00"/> Reference Time: <input type="text" value="14:13:00"/> 5. Acceptance Criteria: within 60 sec. of Reference Time		
SINGLE POINT CHECK				
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Indicated Flow Rate (LPM)	Calibration Flow Rate (Q _a) (LPM)	Percent Difference ⁶ (Sampler vs. Calibration)	Percent Difference ⁷ (Calibration vs. Design)
	16.67	17.22	-3.2%	0.0%
	THREE POINT CHECK			
	Sampler Indicated Flow Rate (LPM)	Calibration Flow Rate (Q _a) (LPM)	Percent Difference ⁶ (Sampler vs. Calibration)	Percent Difference ⁷ (Calibration vs. Design)

A three point flow check is required quarterly and a single point operating flow rate check is required monthly
The 2025i Partisol operates using actual flow rates, and converts to standard conditions in the onboard data file

6. Acceptance Criteria: ± 4.0%
7. Acceptance Criteria: ± 5.0%

Calibration Tech: G. Mazik

**PARTISOL PARTICULATE MATTER (PM) SAMPLER
Calibration Data**

SITE NAME: KCBX ST-NE
AQS Site Code: N/A
DATE: 11-Mar-14
TIME: 13:40

CLIENT: Koch
SAMPLER ID: ST-NE-Carbon
Model Number: 2025i
Serial Number: 20544

TEMPERATURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Sensor Temperature (T _s) (°C)	Calibration Sensor Temperature (T _c) (°C)	Temperature Difference ¹ (Sampler - Calibration) (°C)
	5.7	6.4	-0.7
	5.7	6.4	-0.7
	5.7	6.4	-0.7

1. Acceptance Criteria: ± 2.0 °C

PRESSURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Sensor Pressure (P _s) (mm Hg)	Calibration Sensor Pressure (P _c) (mm Hg)	Pressure Difference ² (Sampler - Calibration) (mm Hg)
	737.0	737.0	0.0
	737.0	737.0	0.0
	737.0	737.0	0.0

2. Acceptance Criteria: ± 10.0 mm Hg

COMPARTMENT TEMPERATURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Compartment Temperature (T _s) (°C)	Calibration Sensor Temperature (T _c) (°C)	Temperature Difference ³ (Sampler - Calibration) (°C)
	7.3	6.9	0.4
	7.3	6.9	0.4
	7.3	6.9	0.4

3. Acceptance Criteria: ± 2.0 °C

LEAK CHECK, TIME, AND FLOWRATE Calibration Data				
External Leak Check ⁴ <input type="text" value="19.0"/> mmHg 4. Acceptance Criteria < 25 mm Hg		Clock/Timer Verification ⁵ : Partisol : <input type="text" value="13:49:00"/> Reference Time: <input type="text" value="13:52:00"/> 5. Acceptance Criteria: within 60 sec. of Reference Time		
SINGLE POINT CHECK				
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Indicated Flow Rate (LPM)	Calibration Flow Rate (Q _a) (LPM)	Percent Difference ⁶ (Sampler vs. Calibration)	Percent Difference ⁷ (Calibration vs. Design)
	16.67	16.65	0.1%	0.0%
	THREE POINT CHECK			
Sampler Indicated Flow Rate (LPM)	Calibration Flow Rate (Q _a) (LPM)	Percent Difference ⁶ (Sampler vs. Calibration)	Percent Difference ⁷ (Calibration vs. Design)	

A three point flow check is required quarterly and a single point operating flow rate check is required monthly
 The 2025i Partisol operates using actual flow rates, and converts to standard conditions in the onboard data file

6. Acceptance Criteria: ± 4.0%
 7. Acceptance Criteria: ± 5.0%

Calibration Tech: G. Mazik

Note: Changed time from 13:49 to 13:52

**PARTISOL PARTICULATE MATTER (PM) SAMPLER
Calibration Data**

SITE NAME: KCBX NT-NE
AQS Site Code: N/A
DATE: 11-Mar-14
TIME: 9:50

CLIENT: Koch
SAMPLER ID: NT-NE-Metals
Model Number: 2025i
Serial Number: 20546

TEMPERATURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Sensor Temperature (T _s) (°C)	Calibration Sensor Temperature (T _c) (°C)	Temperature Difference ¹ (Sampler - Calibration) (°C)
	9.1	9.3	-0.2
	9.1	9.3	-0.2
	9.1	9.3	-0.2

1. Acceptance Criteria: ± 2.0 °C

PRESSURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Sensor Pressure (P _s) (mm Hg)	Calibration Sensor Pressure (P _c) (mm Hg)	Pressure Difference ² (Sampler - Calibration) (mm Hg)
	737.0	738.0	-1.0
	737.0	738.0	-1.0
	737.0	738.0	-1.0

2. Acceptance Criteria: ± 10.0 mm Hg

COMPARTMENT TEMPERATURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Compartment Temperature (T _s) (°C)	Calibration Sensor Temperature (T _c) (°C)	Temperature Difference ³ (Sampler - Calibration) (°C)
	11.1	10.6	0.5
	11.1	10.6	0.5
	11.2	10.6	0.6

3. Acceptance Criteria: ± 2.0 °C

LEAK CHECK, TIME, AND FLOWRATE Calibration Data				
External Leak Check ⁴ <input type="text" value="25.0"/> mmHg 4. Acceptance Criteria < 25 mm Hg		Clock/Timer Verification ⁵ : Partisol : <input type="text" value="9:55:00"/> Reference Time: <input type="text" value="9:57:00"/> 5. Acceptance Criteria: within 60 sec. of Reference Time		
SINGLE POINT CHECK				
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Indicated Flow Rate (LPM)	Calibration Flow Rate (Q _a) (LPM)	Percent Difference ⁶ (Sampler vs. Calibration)	Percent Difference ⁷ (Calibration vs. Design)
	16.68	16.89	-1.2%	0.0%
	THREE POINT CHECK			
	Sampler Indicated Flow Rate (LPM)	Calibration Flow Rate (Q _a) (LPM)	Percent Difference ⁶ (Sampler vs. Calibration)	Percent Difference ⁷ (Calibration vs. Design)

A three point flow check is required quarterly and a single point operating flow rate check is required monthly
 The 2025i Partisol operates using actual flow rates, and converts to standard conditions in the onboard data file

6. Acceptance Criteria: ± 4.0%
 7. Acceptance Criteria: ± 5.0%

Calibration Tech: G. Mazik

Note: Changed time from 09:55 to 09:57.

**PARTISOL PARTICULATE MATTER (PM) SAMPLER
Calibration Data**

SITE NAME: KCBX NT-NE
AQS Site Code: N/A
DATE: 11-Mar-14
TIME: 9:15

CLIENT: Koch
SAMPLER ID: NT-NE-Carbon
Model Number: 2025i
Serial Number: 20545

TEMPERATURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Sensor	Calibration Sensor	Temperature Difference ¹
	Temperature (T _s) (°C)	Temperature (T _c) (°C)	(Sampler - Calibration) (°C)
	9.4	9.5	-0.1
	9.4	9.5	-0.1
9.4	9.5	-0.1	

1. Acceptance Criteria: ± 2.0 °C

PRESSURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Sensor	Calibration Sensor	Pressure Difference ²
	Pressure (P _s) (mm Hg)	Pressure (P _c) (mm Hg)	(Sampler - Calibration) (mm Hg)
	738.0	738.0	0.0
	738.0	738.0	0.0
738.0	738.0	0.0	

2. Acceptance Criteria: ± 10.0 mm Hg

COMPARTMENT TEMPERATURE Calibration Data			
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Compartment	Calibration Sensor	Temperature Difference ³
	Temperature (T _s) (°C)	Temperature (T _c) (°C)	(Sampler - Calibration) (°C)
	11.4	10.6	0.8
	11.4	10.6	0.8
11.4	10.6	0.8	

3. Acceptance Criteria: ± 2.0 °C

LEAK CHECK, TIME, AND FLOWRATE Calibration Data				
External Leak Check ⁴ <input type="text" value="14.0"/> mmHg 4. Acceptance Criteria < 25 mm Hg		Clock/Timer Verification ⁵ : Partisol : <input type="text" value="9:30:00"/> Reference Time: <input type="text" value="9:30:00"/> 5. Acceptance Criteria: within 60 sec. of Reference Time		
SINGLE POINT CHECK				
Calibration Device: BGI Model Number: deltaCal Serial Number: 579 Certification Expiration: 09/10/13	Sampler Indicated	Calibration	Percent Difference ⁶	Percent Difference ⁷
	Flow Rate (LPM)	Flow Rate (Q _a) (LPM)	(Sampler vs. Calibration)	(Calibration vs. Design)
	16.65	16.54	0.7%	0.0%
THREE POINT CHECK				
Sampler Indicated	Calibration	Percent Difference ⁶	Percent Difference ⁷	
Flow Rate (LPM)	Flow Rate (Q _a) (LPM)	(Sampler vs. Calibration)	(Calibration vs. Design)	

A three point flow check is required quarterly and a single point operating flow rate check is required monthly
 The 2025i Partisol operates using actual flow rates, and converts to standard conditions in the onboard data file

6. Acceptance Criteria: ± 4.0%
 7. Acceptance Criteria: ± 5.0%

Calibration Tech: G. Mazik