

Cover Sheet for

ENVIRONMENTAL CHEMISTRY METHOD

Pesticide Name: Thiram

MRID #: 447245-01

Matrix: Water

Analysis: GC/FPD

This method is provided to you by the Environmental Protection Agency's (EPA) Environmental Chemistry Laboratory (ECL). This method *is not* an EPA method but one which was submitted to EPA by the pesticide manufacturer to support product registration. EPA recognizes that the methods may be of some utility to state, tribal, and local authorities, but makes no claim of validity by posting these methods. Although the Agency reviews *all* Environmental Chemistry Methods submitted in support of pesticide registration, the ECL evaluates only about 30% of the currently available methods. Most methods perform satisfactorily but some, particularly the older methods, have deficiencies. Moreover, the print quality of the methods varies considerably because the methods originate from different sources. Therefore, the methods offered represent the best available copies.

If you have difficulties in downloading the method, or further questions concerning the methods, you may contact Elizabeth Flynt at 228-688-2410 or via e-mail at flynt.elizabeth@epa.gov.

MORSE LABORATORIES, INC.

SOP# Meth-95

Revision # 2 Date 07/96

DETERMINATION OF THIRAM IN WATER

Reason for Revision: To provide for the preparation and use of thiram standards as solutions in acetonitrile.

1.0 PRINCIPLE

Thiram present in the sample is converted to CS₂ during reaction with HCl/Stannous chloride reagent at 100°C in a sealed reaction vial. An aliquot of the headspace is injected into a gas chromatograph where the sample responses are compared to a thiram standard similarly prepared and injected. The limit of quantitation for thiram in water is 3.0 µg/L.

2.0 EQUIVALENCE STATEMENT

During the conduct of this analysis, equivalent apparatus, solvents, glassware, or techniques (such as sample concentration) may be substituted for those specified in this method, except where otherwise noted. In the event an equivalent piece of equipment or an equivalent technique is used, its use will be documented in the study records, when appropriate.

3.0 APPARATUS AND EQUIPMENT

Balance: Analytical balance capable of weighing to ±0.1 mg

Reaction vials - 160 mL equipped for crimp sealing with Teflon-lined septums, Pierce Chemical Co., Rockford, IL 61105, cat. #12995 (listed as 125 mL in catalog)

Crimp seals with Teflon-lined septums - Pierce Chemical Company, Rockford, IL 61105
Teflon/Silicone Disc Septums cat #12720
Aluminum seals (20 mm) cat #13214

Crimper - Wheaton Instruments, Millville, NJ, part #224303

SOP# Meth-95, page 2

Water Bath

Gas Chromatograph - Microtek MT-220 or equivalent with a flame photometric detector in the sulfur mode

Gas Chromatographic Column - 6' x 1/4" o.d. x 4 mm i.d. glass column packed with 28% Pennwalt 223 + 4% KOH on 80/100 Gas Chrom R or equivalent

Gas tight syringes - 10, 50, 100, 250, 500, 1000, 2500 μ L, Hamilton Co., Reno, NV

Magnetic Stir Plate with stir bars

"Airpettor" Adjustable pipettes - American SMI "Airpettor", Scientific Products, Sunnyvale, CA 94089 or equivalent

Pipettes: 50-200 μ L volume range cat. #P5086-2
200-1000 μ L volume range cat. #P5086-3

Tips: 2-200 μ L cat. #P5059-37 (Baxter)
100-1000 μ L cat. #P5059-801 (Baxter)

Glass pipettes - Class A graduated, 25 mL

4.0 REAGENTS AND MATERIALS

Thiram: Analytical grade

GLC Column Packing - 28% Pennwalt 223 + 4% KOH on 80/100 Gas Chrom R, Alltech Associates, Inc., Deerfield, IL 60015 or equivalent

Note: The PT 28% Alltech 223 column is equivalent to the Pennwalt 223 column.

EDTA (tetrasodium) - over 98% purity, Mallinckrodt, Paris, Kentucky 40361

SOP# Meth-95, page 3

EDTA Solution -	10% (w/v) in boiled deionized water
Hydrochloric Acid -	concentrated, J.T. Baker Chemical Co., Phillipsburg, N.J. 08665
Stannous Chloride -	analytical grade, EM Science, Gibbstown, NJ 08027
Carbon disulfide -	reagent grade, J.T. Baker Chemical Co. Phillipsburg, N.J. 08665
Methanol -	nanograde, J.T. Baker Chemical Co. Phillipsburg, N.J. 08665
Hexane -	nanograde, J.T. Baker Chemical Co. Phillipsburg, N.J. 08665
Acetonitrile -	Pesticide residue grade
HCl/Stannous chloride reagent -	8N:3%; weigh 3.6 g $\text{SnCl}_2 \cdot (\text{H}_2\text{O})_2$ into a 100 mL volumetric flask. Add 64 mL concentrated HCl. QS to 100 mL with boiled deionized water

5.0 SAMPLE PREPARATION

Samples are quick thawed, using a water bath, at temperatures up to 40°C. Process samples immediately following completion of the thawing procedure.

6.0 STANDARD PREPARATION

6.1 Carbon disulfide (CS₂):

1. 100 μL CS₂ is dissolved in 100 mL hexane.
2. 4 μL of the 0.1% solution is placed in a reaction flask, which is immediately sealed and heated to 100°C to allow all hexane and CS₂ to equilibrate in headspace.

Note: This standard is used to locate the gas chromatographic CS₂ response the first time this analysis is attempted. Thereafter, given similar column and GC conditions, only the active ingredient standard needs to be prepared.

SOP# Meth-95, page 4

6.2 Thiram

Two methods are provided for the preparation of thiram standards used for fortification and instrumental analysis purposes. One produces an aqueous suspension and the other produces a solution. Both methods are suitable for the analysis of field samples and the preparation and analysis of concurrent QC fortification samples. However, only the standard prepared as an aqueous suspension is suitable for use in the preparation of fortified samples intended for stability assessment (field or laboratory).

6.2.1 Preparation of thiram standards as aqueous suspensions

1. Correcting for purity, weigh 10.0 mg active ingredient of a thiram analytical standard into a glass weigh boat, then transfer to a 250 mL beaker.
2. Using a 100 mL volumetric pipet, deliver 100 mL boiled deionized water (cooled) into the beaker containing the thiram standard.
3. Insert a magnetic stir bar in the beaker containing the aqueous suspension and place on a magnetic stir plate. Homogenize the thiram suspension for 10 minutes.

Note: The resulting suspension (100 $\mu\text{g}/\text{mL}$) must be homogeneous and devoid of heavy particles. All subsequent dilutions must also be maintained in a constant suspension by stirring with a magnetic stir bar in a similar fashion as the stock suspension.

4. While the suspension is being homogenized, remove 10.0 mL with a graduated pipet and dilute to 100 mL with boiled deionized water.

Note: 10 mL aliquot must be withdrawn quickly and accurately on first attempt. If the initial withdrawal overshoots or undershoots the 10 mL mark, the withdrawal procedure must be reinitiated. This produces a 10 $\mu\text{g}/\text{mL}$ solution.

5. An instrumentation standard with a 0.1 $\mu\text{g}/\text{mL}$ headspace concentration is prepared by adding 1.35 mL of the 10 $\mu\text{g}/\text{mL}$ suspension (using an Airpette) to a 160 mL vial.

Note: Two 675 μL aliquots can be used in place of 1.35 mL.

SOP# Meth-95, page 5

6. Add 8.65 mL 10% EDTA solution and 15 mL HCl/Stannous chloride reagent to the reaction vial and immediately crimp seal.
7. React contents of vial in the same fashion as samples, as discussed later. Use this standard for gas chromatography.

Note: Based on the stability of thiram in water suspension (<3% degradation in 30 minutes), all manipulations with suspension standards must be completed within this time frame.

6.2.2 Preparation of thiram standards as acetonitrile solutions

1. Correcting for purity, weigh 10.0 mg active ingredient of a thiram analytical standard into a glass weigh boat, then transfer to a 100 mL volumetric flask.
2. Bring to volume with acetonitrile and mix to dissolve. This produces a 100 $\mu\text{g}/\text{mL}$ solution.
3. Transfer 10.0 mL of the 100 $\mu\text{g}/\text{mL}$ solution with a graduated pipet to a 50 mL volumetric flask. Bring to volume with acetonitrile. Mix well. This produces a 20 $\mu\text{g}/\text{mL}$ solution.
4. An instrumentation standard with a 0.1 $\mu\text{g}/\text{mL}$ headspace concentration is prepared by adding 675 μL of the 20 $\mu\text{g}/\text{mL}$ suspension to a 160 mL reaction vial.
5. Add 10.0 mL 10% EDTA solution and 15 mL HCl/Stannous chloride reagent to the reaction vial and immediately crimp seal.
6. React contents of vial in the same fashion as samples, as discussed later. Use this standard for gas chromatography.

Both standard preparations detailed above are intended for analyses with a limit of quantitation of 3.0 $\mu\text{g}/\text{L}$. Instrumentation standards of differing concentrations may be prepared to enable the analysis of samples with limits of quantitation other than 3.0 $\mu\text{g}/\text{L}$.

SOP# Meth-95, page 6.

7.0 SAMPLE FORTIFICATION

1. Spike samples are fortified at the correct level by adding the appropriate volume from either the 1.0 $\mu\text{g/mL}$, 10 $\mu\text{g/mL}$, 20 $\mu\text{g/mL}$ or 100 $\mu\text{g/mL}$ fortification suspension or solution as applicable.

Note: Regarding *suspension* standards, 1.0 $\mu\text{g/mL}$ is only prepared if necessary. While the 10 $\mu\text{g/mL}$ *suspension* is being homogenized, remove 10.0 mL and dilute to 100 mL with boiled deionized water.

2. When using *suspension* standards only pipettes or "Airpette" should be employed to add the fortification solutions to the sample. **DO NOT USE MICROLITER SYRINGES.**
3. Using a 25 mL graduated pipet, place 20.0 mL of thawed water sample into a 160 mL reaction vial.
4. Fortify the spike samples with the appropriate concentration. When using *suspensions*, aliquots are removed only while the suspension is being actively homogenized.

Use volumes of no greater than 1.0 mL when fortifying with *solution* standards.

5. Add 10% EDTA solution to make a total volume of 30.0 mL (sample plus fortification solution plus 10% EDTA solution; inclusion of fortification solution volume only applies to use of *suspension* standards). See Discussion, Section 11.0.
6. Add 15.0 mL HCl/Stannous Chloride (8N/3%) reagent. Immediately seal.
7. Treat as samples from this point on, heating the fortification at 100°C for the same time as the samples.

8.0 SAMPLE EXTRACTION

1. Using a 25 mL graduated pipet, place 20.0 mL of thawed water sample into a 160 mL reaction vial.
2. Add 10% EDTA solution to make a total volume of 30.0 mL (sample plus 10% EDTA solution). Volume of 10% EDTA added is typically 10.0 mL.

SOP# Meth-95, page 7

3. Add 15.0 mL HCl/Stannous Chloride (8N/3%) reagent. Immediately seal the reaction vial.
4. Place vial in boiling waterbath for a total of 2 hours, hand-shaking vials approximately every 5 minutes for the first 30 minutes, then every 30 minutes for the remaining 1 1/2 hours.
5. After reaction, maintain sample at 100°C in the waterbath during GLC analysis.

9.0 GAS CHROMATOGRAPHIC ANALYSIS

- 9.1 The following column and conditions are suggested. Other columns that exhibit proper analyte chromatography and resolution from interferences may be used when necessary.

Column: 6' x 1/4" o.d. glass column packed with 28% Pennwalt 223 + 4% KOH on 80/100 mesh Gas Chrom R

Carrier Gas: Nitrogen at 45 mL/min.

Temperatures: Injector: 210 °C
Detector: 160 °C
Column: 125 °C isothermal

Injection Volume: Variable. Inject 2-1000 µL of airspace from samples into GC using airtight syringes. Compare sample responses to those produced in the standard curve.

Retention Time: ~2.0 minutes

9.2 Calibration

1. Prepare a 4-point standard curve by injecting 0.3 ng through 2.5 ng (3 µL through 25 µL of a 0.1 µg/mL headspace concentration standard) of thiram.

Note: 2.5 ng may produce an off-scale response on some FPD detectors due to their inherent differences in logarithmic response. In such cases, inject an amount of standard which produces approximately a 90% full scale deflection (FSD) as the high point of the curve.

SOP# Meth-95, page 8

2. In any case, 0.3 ng, representing a 2.16 $\mu\text{g/L}$ response when 800 μL of sample is injected, must be included as the lowest point on the standard curve and should produce a response peak height of at least 5 mm. Identification of the peak produced by the thiram standard as CS_2 is achieved by demonstrating its retention time to be identical to that of the peak produced by an injection of CS_2 standard. This identification process needs to be conducted only if the CS_2 retention time needs to be identified (i.e., when a new GC column is employed or instrument repairs which might have affected retention time have occurred).

Compare sample responses to those produced in the standard curve.

9.3 Notes

1. Condition column overnight at 200 °C with 30 mL/minute carrier gas flow.
2. Prior to analysis, make several injections (10-100 μL) of the thiram standard (for example, 0.1 $\mu\text{g/mL}$ headspace as appropriate) to sensitize the column for CS_2 .
3. The column and conditions stated in the method have been satisfactory for the matrix being analyzed. The specific column packing/coating, carrier gas, column temperature and flow rate listed are typical conditions for this analysis. Specific conditions used will be noted on each chromatographic run and will not otherwise be documented.

10.0 CALCULATIONS

Calculations from the instrumental analysis are conducted using a validated software application to generate a standard curve based on nonlinear regression. The data points for analyte concentration (ng) and peak response (mm) are obtained from strip chart recorder-generated chromatograms. All standards injected for the standard curve and their corresponding peak height responses are entered into the program.

A power series equation is used to determine unknown x values when given the y value. The equation that defines the curve produced when a logarithmic relationship exists between concentration and peak responses is:

$$y = ax^b$$

SOP# Meth-95, page 9

where:

y = peak height response (mm)

x = nanogram found for the peak of interest

a and b = variables dependent on data points entered

The calculations for $\mu\text{g/L}$ found and percent recovery (for fortified samples) are:

- The amount of Thiram (in $\mu\text{g/L}$) found in the sample is calculated according to the following equation:

$$\mu\text{g/L thiram} = \text{ng found} \times \frac{\text{final vol. (mL)}}{\text{sample vol. (mL)}} \times \frac{1}{\text{inj. vol. } (\mu\text{L})} \times \frac{1000 \mu\text{L}}{1 \text{ mL}} \times \frac{1 \mu\text{g}}{1000 \text{ ng}} \times \frac{1000 \text{ mL}}{1 \text{ L}}$$

where:

ng found = ng found calculated via a software program

final volume (mL) = headspace volume of sample-containing reaction vial (115 mL)

sample vol. (mL) = volume of sample analyzed in mL

injection volume (μL) = μL amount of sample injected into gas chromatograph

1000 $\mu\text{L/mL}$ = conversion factor

1 $\mu\text{g}/1000 \text{ ng}$ = conversion factor

1000 mL/L = conversion factor

SOP# Meth-95, page 10

2. The percent recovery for fortified control samples is calculated as follows:

$$\% \text{ Recovery} = \frac{\mu\text{g/L found in fortified control} - \mu\text{g/L found in control}}{\mu\text{g/L added}} \times 100$$

11.0 DISCUSSION

- The gas chromatographic method described herein has a limit of quantitation of 3.0 $\mu\text{g/L}$ for thiram.
- In order to eliminate a possibility of CS_2 contamination, reaction vials were rinsed with methanol and baked at 135°C for 45 minutes. Teflon-lined septums were rinsed with methanol and air dried. The use of rubber or plastic utensils was avoided.

The use of Teflon-lined rubber septums is acceptable for temporary storage of rinsed, baked reaction vials.

- Standard curves are prepared by injection of variable volumes of a single standard preparation at a concentration of 0.1 $\mu\text{g/mL}$ headspace (or other headspace concentrations, as applicable).
- Samples must be kept frozen at all times until quick-thawed for removal of sub-sample for extraction.
- Samples must be reacted in a timely manner following addition of reaction reagents. Once reacted, the samples (now containing in the form of CS_2 any thiram that may have been present) may be stored at room temperature overnight. Simply reheat the samples the following day at 100°C for approximately 30 minutes with shaking.
- When the volume of fortification solution added to the reaction vial is less than 1.0 mL, the volume of 10% EDTA added to obtain a final headspace volume of 115 mL is not adjusted to account for the addition of the fortification solution.

The addition of less than 1.0 mL of solution would account for at most, 0.87% of the total volume of headspace. It is felt by Morse Laboratories, Inc. that this percentage is insignificant and subtracting the fortification solution volume from the volume of 10% EDTA to be added is not necessary.



1525 Fulton Avenue
Sacramento, California 95825
TEL: 916-481-3141
FAX: 916-481-2959

December 13, 1996

Method Modifications by Morse Laboratories, Inc. to
Method SOP, Meth-95; Revision #2

Method Reference: "Determination of Thiram in Water", (Meth-95, Revision #2, 07/96)

Modifications to Method Sections:

6.2.1 Preparation of Thiram Standards as Aqueous Suspensions

Step 5: An instrumentation standard with a 0.1 $\mu\text{g/mL}$ headspace concentration is prepared by adding 1.15 mL of the 10 $\mu\text{g/mL}$ suspension (using an airpette) to a 160 mL vial:

Note: Two 575 μL aliquots can be used in place of 1.15 mL.

Step 6: Add 8.85 mL 10% EDTA solution and 15 mL HCl/stannous chloride reagent to the reaction vial and immediately crimp seal.

Reason for Modifications: To standardize the headspace volume to 115 mL for the samples being analyzed and the instrumentation standard prepared for analysis of those samples.

???

1.15
8.85

15
25.00

160 - 25 = 135

Head space vol. for standard is still 135 mL