



Site 42-017-0012: Bristol, PA

Met One BAM-1020 Mass Monitor w/VSCC

Continuous (y) vs. FRM (x) $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




FRM minus Continuous $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




Mean $PM_{2.5}$ ($\mu g \cdot m^{-3}$)

#	FRM	Con.	Ratio (Con/FRM)
○ 29	11.6	13.8	1.19
△ 10	14.6	16.1	1.1
▽ 19	10	12.5	1.25
0 29	11.6	13.8	1.19

Additive (y) vs. Multiplicative (x) Bias



R (y) vs. FRM CCV (x)



○ AllData △ Autumn ▽ Spring △ Summer ▽ Winter
8 2008 9 2009 0 2010