



Site 33-015-0014: Portsmouth, NH

Met One BAM-1020 Mass Monitor w/VSCC

Continuous (y) vs. FRM (x) $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




FRM minus Continuous $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




Mean $PM_{2.5}$ ($\mu g \cdot m^{-3}$)

	#	FRM	Con.	Ratio (Con/FRM)
○	39	7.3	7.8	1.07
▽	10	5.3	6	1.13
△	3	10.5	9.5	0.9
▽	26	7.7	8.3	1.08
○	39	7.3	7.8	1.07

Additive (y) vs. Multiplicative (x) Bias



R (y) vs. FRM CCV (x)



○ AllData △ Autumn ▽ Spring △ Summer ▽ Winter
8 2008 9 2009 0 2010