



Site 39-153-0017: Akron, OH

Met One BAM-1020 Mass Monitor w/VSCC

Continuous (y) vs. FRM (x) $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




FRM minus Continuous $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




Mean $PM_{2.5}$ ($\mu g \cdot m^{-3}$)

#	FRM	Con.	Ratio (Con/FRM)
○ 162	13.4	16.2	1.21
▽ 18	16.4	19.2	1.17
△ 45	16.8	20.1	1.2
▽ 99	11.4	13.8	1.21
0 162	13.4	16.2	1.21

Additive (y) vs. Multiplicative (x) Bias



R (y) vs. FRM CCV (x)



○ AllData △ Autumn ▽ Spring △ Summer ▽ Winter
8 2008 9 2009 0 2010