



Site 18-097-0078: Indianapolis, IN

Thermo Scientific 8500C FDMS w/VSCC

Continuous (y) vs. FRM (x) $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




FRM minus Continuous $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




Mean $PM_{2.5}$ ($\mu g \cdot m^{-3}$)

#	FRM	Con.	Ratio (Con/FRM)
○ 172	13	16.3	1.25
▽ 75	15.6	17.7	1.13
△ 5	12.9	14.6	1.13
▽ 92	10.8	15.3	1.42
0 172	13	16.3	1.25

Additive (y) vs. Multiplicative (x) Bias



R (y) vs. FRM CCV (x)



○ AllData △ Autumn ▽ Spring △ Summer ▽ Winter
8 2008 9 2009 0 2010