



Site 18-097-0081: Indianapolis (Remainder), IN

Thermo Scientific 8500C FDMS w/VSCC

Continuous (y) vs. FRM (x) $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




FRM minus Continuous $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




Mean $PM_{2.5}$ ($\mu g \cdot m^{-3}$)

#	FRM	Con.	Ratio (Con/FRM)
○ 303	13.4	13.6	1.01
▽ 103	16.3	14.8	0.91
△ 81	13	14	1.08
△ 11	13.3	15.7	1.18
▽ 108	11.1	11.9	1.07
9 94	13.1	13.8	1.05
0 209	13.6	13.5	0.99

Additive (y) vs. Multiplicative (x) Bias



R (y) vs. FRM CCV (x)



○ AllData △ Autumn ▽ Spring △ Summer ▽ Winter
8 2008 9 2009 0 2010