

US EPA ARCHIVE DOCUMENT

**ATTACHMENT A (revised)**

Table 1: Evaluation of Performance Evaluation Sample Results for Cadmium

Sample ID	Laboratory Results -- Cadmium ppb				Evaluation Criteria	
	KHF	Result Evaluation	EPA	Result Evaluation	True Value	Acceptable Result Range
KH1	-40	<b>Fail</b>	2.5	<b>Pass</b>	0	-11 to 11
KH2	78	<b>Fail</b>	94	<b>Pass</b>	100	80 to 120
KH3	178	<b>Pass</b>	180	<b>Pass</b>	200	160 to 240
KH4	178	<b>Pass</b>	180	<b>Pass</b>	200	160 to 140
KH5	289	<b>Pass</b>	270	<b>Pass</b>	300	240 to 360

Table 2: Evaluation of Performance Evaluation Sample Results for Zinc

Sample ID	Laboratory Results -- Zinc ppm				Evaluation Criteria	
	KHF	Result Evaluation	EPA	Result Evaluation	True Value	Acceptable Result Range
KH9	-0.03	<b>Pass</b>	NA	NA	0	-1 to 1
KH13	204	<b>Fail</b>	710	<b>Pass</b>	700	560 to 840
KH14	221	<b>Fail</b>	1100	<b>Pass</b>	1100	880 to 1320
KH15	232	<b>Fail</b>	1400	<b>Pass</b>	1500	1200 to 1800
KH16	233	<b>Fail</b>	NA	NA	1500	1200 to 1800
KH17	243	<b>Fail</b>	NA	NA	1900	1520 to 2280